



Agilent Technologies

DECLARATION OF CONFORMITY

According to EN ISO/IEC 17050-1:2010



Manufacturer's Name: Agilent Technologies Australia (M) Pty Ltd
Manufacturer's Address: 679 Springvale Rd,
Mulgrave, Victoria, 3170
Australia

Declares under sole responsibility that the product as originally delivered

Product Name: Temperature probe
Model Number: G9812A
Product Options: This declaration covers all options of the above product

complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

EMC Directive 2014/30/EU
RoHS Directive 2011/65/EU

and conforms with the following product standards:

	Standard	Limit
EMC:	IEC 61326-1:2005 CISPR 11:2010 Canada: ICES-001 / NMB-001 Issue 4, June 2006	Group 1 Class A
Other :	EN 50581	

Supplementary Information:

The products were tested in a typical configuration with Agilent Technologies test systems.

This DoC applies to above-listed products placed on the EU market after:

June 1, 2017

Date

Cale Hunkin, Quality Manager

For further information, please contact your local Agilent Technologies sales office, agent or distributor.
or Agilent Technologies Deutschland GmbH, Hewlett-Packard-Strasse 8, 76337 Waldbronn, Germany.

Part Number:



8510145200

Revision: 5