



Agilent Technologies

DECLARATION OF CONFORMITY

According to EN ISO/IEC 17050-1:2010



Manufacturer's Name: Agilent Technologies Australia (M) Pty Ltd
Manufacturer's Address: 679 Springvale Rd,
Mulgrave, Victoria, 3170
Australia

Declares under sole responsibility that the product as originally delivered

Product Name: Cary Eclipse
Model Number: G9800A, G9801A, G9802A, G9803A
Product Options: This declaration covers all options of the above product(s)

complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

Low Voltage Directive 2014/35/EU
EMC Directive 2014/30/EU
RoHS Directive 2011/65/EU

and conforms with the following product standards:

Standard	Limit
EMC: IEC 61326-1:2012 / EN 61326-1:2013 Australia/New Zealand AS/NZS CISPR 11:2004 Canada: ICES-001 / NMB-001 Issue 4, June 2006	Class 1 Group A

Safety: IEC 61010-1:2010 / EN 61010-1:2010
IEC 61010-2-081:2015 / EN 61010-1:2015
Canada: CAN/CSA-C22.2 No. 61010-1-12
Canada: CAN/CSA-C22.2 No. 61010-2-081-15
USA: ANSI/UL 61010-1-2012
USA: UL 61010-2-081:2015

Other: EN 50581:2012

Supplementary Information:

The products were tested in a typical configuration with Agilent Technologies test systems.

This DoC applies to above-listed products placed on the EU market after:

December 18, 2017

Date

Cale Hunkin, Quality Manager

For further information, please contact your local Agilent Technologies sales office, agent or distributor or Agilent Technologies Deutschland GmbH, Hewlett-Packard-Strasse 8, 76337 Waldbronn, Germany.

Part Number:



8510178400

Revision: 9