



**Agilent Technologies**

**DECLARATION OF CONFORMITY**

According to EN ISO/IEC 17050-1:2010



**Manufacturer's Name:** Agilent Technologies Australia (M) Pty Ltd

**Manufacturer's Address:** 679 Springvale Rd,  
Mulgrave, Victoria, 3170  
Australia

**Declares under sole responsibility that the product as originally delivered**

**Product Name:** Microplate reader

**Model Number:** G9810A

**Product Options:** This declaration covers all options of the above product(s)

**complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:**

Low Voltage Directive 2014/35/EU

EMC Directive 2014/30/EU

RoHS Directive 2011/65/EU

**and conforms with the following product standards:**

	<b>Standard</b>	<b>Limit</b>
<b>EMC</b>	IEC 61326-1:2005 Ed 1 Australia/New Zealand AS/NZS CISPR11:2004 Canada: ICES / NMB-001: Issue 4, June 2006	Group 1 Class A
<b>Safety:</b>	IEC 61010-1: 2001 IEC 61010-2-081:2001 + A1:2003	
<b>Other:</b>	EN 50581	

**Supplementary Information:**

The products were tested in a typical configuration with Agilent Technologies test systems.

**This DoC applies to above-listed products placed on the EU market after:**

June 1, 2017


Date

**Cale Hunkin,**

Quality Manager

For further information, please contact your local Agilent Technologies sales office, agent or distributor or Agilent Technologies Deutschland GmbH, Hewlett-Packard-Strasse 8, 76337 Waldbronn, Germany.

Part Number:



8510178700

Revision: 7