



**Agilent Technologies**

**DECLARATION OF CONFORMITY**

According to EN ISO/IEC 17050-1:2004



**Manufacturer's Name:** Agilent Technologies Australia (M) Pty Ltd  
**Manufacturer's Address:** 679 Springvale Rd,  
Mulgrave, Victoria, 3170  
Australia

**Declares under sole responsibility that the product as originally delivered**

**Product Name:** Multicell Holder 4 Position  
**Model Number:** G9808A  
**Product Options:** This declaration covers all options of the above product(s)

**complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:**

Low Voltage Directive 2014/35/EU  
EMC Directive 2014/30/EU  
RoHS Directive 2011/65/EU

**and conforms with the following product standards:**

	<b>Standard</b>	<b>Limit</b>
<b>EMC:</b>	IEC 61326-1:2012 / EN 61326-1:2013 Australia/New Zealand AS/NZS CISPR 11:2004 Canada: ICES / NMB-001: Issue 4, June 2006	Group 1 Class A
<b>Safety:</b>	IEC 61010-1:2001 IEC 61010-2-081:2001 + A1:2003	
<b>Other:</b>	EN 50581	

**Supplementary Information:**

The products were tested in a typical configuration with Agilent Technologies test systems.

**This DoC applies to above-listed products placed on the EU market after:**

June 1, 2017

Date

**Cale Hunkin, Quality Manager**

For further information, please contact your local Agilent Technologies sales office, agent or distributor.  
or Agilent Technologies Deutschland GmbH, Hewlett-Packard-Strasse 8, 76337 Waldbronn, Germany.



Part Number: 8510178800

Revision: 8