



**Agilent Technologies**

**DECLARATION OF CONFORMITY**

According to EN ISO/IEC 17050-1:2010



**Manufacturer's Name:** Agilent Technologies Australia (M) Pty Ltd  
**Manufacturer's Address:** 679 Springvale Rd  
Mulgrave, VIC, 3170  
Australia

**Declares under sole responsibility that the product as originally delivered**

**Product Name:** Cary 100 Bio Spectrophotometer  
Cary 100 Bio Spectrophotometer & ACB  
Cary 300 Bio Spectrophotometer  
Cary 300 Bio Spectrophotometer & ACB  
**Model Number:** G9820A, G9821A, G9822A, G9823A  
**Product Options:** This declaration covers all options of the above products

**complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:**

Low Voltage Directive 2014/35/EU  
EMC Directive 2014/30/EU  
RoHS Directive 2011/65/EU

**and conforms with the following product standards:**

	<b>Standard</b>	<b>Limit</b>
<b>EMC:</b>	IEC 61326-1:2012 / EN 61326:2013 Australia/New Zealand AS/NZS CISPR 11:2004 Canada: ICES / NMB-001 Issue 4, June 2006	Group 1 Class A
<b>Safety:</b>	IEC 61010-1:2010 / EN 61010-1:2010 IEC 61010-2-081:2015 / EN 61010-2-081:2015 Canada: CAN/CSA-C22.2 No. 61010-1-15 Canada: CAN/CSA-C22.2 No. 61010-2-081-15 USA: ANSI/UL 61010-1-2012 USA: ANSI/UL 61010-2-081:15	
<b>Other:</b>	EN 50581:2012	

**Supplementary Information:**

The products were tested in a typical configuration with Agilent Technologies test systems  
**This DoC applies to above-listed products placed on the EU market after:**

December 18, 2017

Date

**Cale Hunkin, Quality Manager**

For further information, please contact your local Agilent Technologies sales office, agent or distributor,  
or Agilent Technologies Deutschland GmbH, Hewlett-Packard-Strasse 8, 76337 Waldbronn, Germany.

Part Number:



8510189800

Revision: 8