



Agilent Technologies

DECLARATION OF CONFORMITY

According to EN ISO/IEC 17050-1:2010



Manufacturer's Name: Agilent Technologies Australia (M) Pty Ltd
Manufacturer's Address: 679 Springvale Rd
Mulgrave, VIC, 3170
Australia

Declares under sole responsibility that the product as originally delivered

Product Name: Cary 4000, Cary 5000, Cary 6000i, Cary 7000
Model Number: G9824A, G9825A, G9826A, G6873A
Product Options: This declaration covers all options of the above products

complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

Low Voltage Directive 2014/35/EU
EMC Directive 2014/30/EU
RoHS Directive 2011/65/EU

and conforms with the following product standards:

	Standard	Limit
EMC:	IEC 61326-1:2012 / EN 61326-1:2013 CISRP 11 / EN 55011 Australia/New Zealand AS/NZS CISPR 11:2010 Canada: ICES-001:2014 / NMC-001:2004	Group 1 Class A
Safety:	IEC 61010-1:2010 / EN 61010-1:2010 IEC 61010-2-081:2015 / EN 61010-2-081:2015 Canada: CAN/CSA-C22.2 No. 61010-1-12 CAN/CSA-C22.2 No. 61010-2-081-04 USA: ANSI/UL 61010-1-2012 USA: ANSI/UL 61010-2-081:15	
Other:	EN 50581:2012	

Supplementary Information:

The products were tested in a typical configuration with Agilent Technologies test systems.

This DoC applies to above-listed products placed on the EU market after:

December 18, 2017

Date

Cale Hunkin, Quality Manager

For further information, please contact your local Agilent Technologies sales office, agent or distributor,
or Agilent Technologies Deutschland GmbH, Hewlett-Packard-Strasse 8, 76337 Waldbronn, Germany.

Part Number:



8510196600

Revision: 9