



Agilent Technologies

DECLARATION OF CONFORMITY
According to ISO/IEC Guide 22 and CEN/CENELEC EN 45014



Manufacturer's Name: Agilent Technologies Australia (M) Pty Ltd
Manufacturer's Address: 679 Springvale Rd,
Mulgrave, Victoria, 3170
Australia

Declares under sole responsibility that the product as originally delivered

Product Name: Agilent 325 UV/VIS Dual Wavelength Detector
Model Number: G9309A
Product Options: This declaration covers all options of the above products

complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

Low Voltage Directive 2006/95/EC
EMC Directive 2004/108/EC

and conforms with the following product standards:

EMC	Standard	Limit
	IEC 61326-1 :2005 / EN 61326-1:2006 CISPR 11 / EN 55011 IEC 61000-3-2 / EN61000-3-2 IEC 61000-3-3 / EN61000-3-3 IEC 61000-4-2 / EN 61000-4-2 IEC 61000-4-3 / EN 61000-4-3 IEC 61000-4-4 / EN 61000-4-4 IEC 61000-4-5 / EN 61000-4-5 IEC 61000-4-6 / EN 61000-4-6 IEC 61000-4-11 / EN 61000-4-11 Canada ICES / NMB-001:2004 Australia/New Zealand: AS/NZS CISPR:2004	Group 1 Class A 4kV CD, 4kV AD 3 V/m: 80-1000 MHz, 1.4-2GHz; 1V/m: 2-2.7GHz 0.5 kV signal lines, 1 kV power lines 0.5 kV line-line, 1 kV line-ground 3Vrms, 0.15-80 MHz 240V~: 0%: half cycle, 0%: 1 cycle, 70%: 25/30 cycles
Safety	IEC 61010-1:2001 / EN 61010-1:2001 Canada: CAN/CSA C22.2 No.61010.1-04 USA: ANSI/UL 61010-1:2001	

Supplementary Information:

The product was tested in a typical configuration.
This ISM device complies with Canadian ICES-001. Cet appareil ISM est conforme à la norme NMB-001 du Canada.

This DoC applies to above-listed products placed on the EU market after:

February 21, 2012

Date

Shunichiro Cho,
Quality Manager

For further information, please contact your local Agilent Technologies sales office, agent or distributor,
or Agilent Technologies Deutschland GmbH, Herrenberger Straße 130, 71034 Böblingen, Germany.

