

**Agilent Technologies****DECLARATION OF CONFORMITY**

According to EN ISO/IEC 17050-1:2010



Manufacturer's Name: Agilent Technologies Singapore (International) Pte., Ltd.
Manufacturer's Address: No. 1 Yishun Ave 7
Supplier's Address: Singapore 768923

Declares under sole responsibility that the product as originally delivered

Product Name: SureScan Microarray Scanner
Model Number: G2600D
Product Options: This declaration covers all options of the above product(s)

complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

Low Voltage Directive 2014/35/EU
 EMC Directive 2014/30/EU

and conforms with the following product standards (note: only applicable tests are shown):

Safety IEC 61010-1:2010 / EN 61010-1:2010

Canada: CAN/CSA-C22.2 No 61010-1-12
 USA: ANSI/UL 61010-1 (3rd edition)

EMC

Standard
 IEC 61326-1:2012 / EN 61326-1:2013

Limit
 Basic EM Environment

Reference Standards:

RFI Emissions:	CISPR 11:2009 + A1:2010 / EN 55011:2009 + A1:2010	Group 1, Class A
ESD:	IEC 61000-4-2:2008 / EN 61000-4-2:2009	4 kV/8 kV contact/air
Radiated RFI Immunity:	IEC 61000-4-3:2006 + A1:2007 + A2:2010 / EN 61000-4-3:2006 + A1:2008 + A2:2010	3 V/m (80 MHz-1 GHz); 3 V/m (1.4-2 GHz); 1 V/m (2- 2.7 GHz)
EFT:	IEC 61000-4-4:2004 + A1:2010 / EN 61000-4-4:2004 + A1:2010	0.5 kV signal lines, 1 kV power lines
Surge:	IEC 61000-4-5:2005 / EN 61000-4-5:2006	0.5 kV line-line, 1 kV line-ground
Conducted RFI Immunity:	IEC 61000-4-6:2008 / EN 61000-4-6:2009	3 V (150 kHz - 80 MHz)
Magnetic Immunity:	IEC 61000-4-8:2009 / EN 61000-4-8:2010	3 A/m, 50 & 60 Hz
Voltage Dips:	IEC 61000-4-11:2004 / EN 61000-4-11:2004	0% - 1/2 cycle, 0% - 1 cycle, 70% - 0.5 sec
Voltage Interruptions:	IEC 61000-4-11:2004 / EN 61000-4-11:2004	70% - 0.5 sec, 0% - 5 sec.

Canada: ICES/NMB-001:2014
 Australia/New Zealand: AS/NZS CISPR 11:2010

Supplementary Information:

The product was tested in a typical configuration with Agilent Technologies test systems.

This DoC applies to above-listed products placed on the EU market after:

20 April 2016

Date

Alka Vyas
 Alka Vyas

DGG QA/RA Manager

For further information, please contact your local Agilent Technologies sales office, agent or distributor, or Agilent Technologies Deutschland GmbH, Hewlett-Packard-Strasse 8, 76337 Waldbronn, Germany.