



**Agilent Technologies**

**DECLARATION OF CONFORMITY**

According to EN ISO/IEC 17050-1:2004



**Manufacturer's Name:** Agilent Technologies Singapore (International) Pte. Ltd.  
**Manufacturer's Address:** No. 1 Yishun Ave 7  
SINGAPORE 768923  
Singapore

**Declares under sole responsibility that the product as originally delivered**

**Product Name:** MassCode UV Module  
**Model Number:** G5201A  
**Product Options:** This declaration covers all options of the above product(s)

**complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:**

EMC Directive (2004/108/EC)  
Low Voltage Directive (2006/95/EC)

**and conforms with the following product standards (note: only applicable tests are shown):**

**Safety** IEC 61010-1:2001 / EN 61010-1:2001  
Canada: CSA C22.2 No. 61010-1-04  
USA: ANSI/UL 61010-1: 2004  
Australia/New Zealand: AS 61010.1-2003

**EMC** Standard  
IEC 61326-1:2005/ EN 61326-1:2006

Limit

*Reference Standards*

RFI Emissions:	CISPR 11:2004 / EN 55011:2007	Group 1, Class A
ESD:	IEC 61000-4-2:2001 / EN 61000-4-2:2001	4 kV/4 kV contact/air
Radiated RFI:	IEC 61000-4-3:2002 / EN 61000-4-3:2002	3 V/m- 80MHz-1GHz/1.4-2GHz, 1V/m- 2- 2.7 GHz
EFT:	IEC 61000-4-4:2004 / EN 61000-4-4:2004	0.5 kV signal lines, 1 kV power lines
Surge:	IEC 61000-4-5:2005 / EN 61000-4-5:2006	0.5 kV line-line, 1 kV line-ground
Conducted RFI:	IEC 61000-4-6:2008 / EN 61000-4-6:2001	3 V, 150kHz-80MHz
Voltage Dips:	IEC 61000-4-11:2004 / EN 61000-4-11:2004	0% - 1/2 cycle, 0% - 1 cycle
Voltage Interrupt:	IEC 61000-4-11:2004 / EN 61000-4-11:2004	70% - 0.5 sec, 0% - 5 sec.
Canada:	ICES/NMB-001:2004	
Australia/New Zealand:	AS/NZS CISPR 11:2004	

The product was tested in a typical configuration with Agilent Technologies test systems.

**Supplementary Information:**

This device was tested in a representative system.

**This DoC applies to above-listed products placed on the EU market after:**

29 April 2011

Date

Name **PHILIP KLIMBAL**

Title **PROGRAM MANAGER**

For further information, please contact your local Agilent Technologies sales office, agent or distributor,  
or Agilent Technologies Deutschland GmbH, Herrenberger Straße 130, 71034 Böblingen, Germany.

Doc No. G5201-90501

Revision A.00