



Manufacturer's Name: Agilent Technologies, Inc.
Manufacturer's Address: 5301 Stevens Creek Boulevard
Santa Clara, California 95051
USA

Declares under sole responsibility that the product as originally delivered

Product Name & Function: RapidFire 365 High-throughput MS System
Model & Part Numbers: Model G9530AA
Product Options: This declaration covers all options of the above products
Serial Number(s) SGYYWWNNNN (YY=13-99, WW=01-53, NNNN=0000-9999)

complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

EMC Directive 2004/108/EC
Machinery Directive 2006/42/EC

and conforms with the following product standards:

EMC: IEC 61326-1:2005 / EN 61326-1:2006
Canada ICES / NMB-001:2004
Australia/New Zealand: AS/NZS CISPR:2004
Safety: IEC 61010-1:2001 / EN 61010-1:2001
IEC 61010-2-081:2009 / EN 61010-2-081:2002/A1:2003
Canada: CAN/CSA-C22.2 No. 61010-1-04
CAN/CSA-C22.2 No. 61010-2-081-04
USA: UL Std. No. 61010-1 (2nd Edition)

Contact established in the Community authorized to compile the technical file or the relevant technical documents:

Agilent Technologies Deutschland GmbH
Herrenbergerstrasse 130
71034 Boeblingen
Germany

Supplementary Information:

The products were tested in a typical configuration with Agilent Technologies test systems
This DoC applies to above-listed products placed on the EU market after:

13-DEC-2013
Date

Santa Clara, California USA
Location


Larry Durandette, Quality Manager

For more information, please contact your local Agilent Technologies sales office, agent or distributor.